



<b>Product Name</b>	FabRICATOR® MagIC
<b>Product Number</b>	A0-FRM-024, A0-FRM-096
<b>Product Lot Number</b>	A0FRM-13G02B
<b>Product Composition</b>	Cross-linked magnetic agarose beads with FabRICATOR®
<b>Manufacturer</b>	Genovis AB
<b>Storage Temperature</b>	+4-8°C

TEST	SPECIFICATIONS	RESULTS LOT
<b>Digestion capacity by SDS-page</b>	≥ 95% of 20 and 200 µg hIgG/20 µl resin in 100 µl 10mM Sodium Phosphate, 150mM NaCl pH 7.4 at 37°C in 20 min	Pass
<b>Bioburden*</b>	0 cfu/100 µl 25% slurry	Pass
<b>Production Date</b>		July 2020

\* Tested on Dextrose Agar, Fluid Thioglycollate Medium and Blood Agar.

**Anna Åkesson**  
Senior QC Engineer  
Genovis AB, Lund, SWEDEN