

# Certificate of Analysis



GENOVIS

Issued April 2018

<b>Product Name</b>	FragIT™ Midispin
<b>Product Number</b>	A0-FR6-100
<b>Product Lot Number</b>	A0FR6-11K14B
<b>Product Composition</b>	Highly cross-linked agarose resin with FabRICATOR®.
<b>Manufacturer</b>	Genovis AB
<b>Storage Temperature</b>	+4-8°C

TEST	SPECIFICATIONS	RESULTS LOT
<b>Digestion capacity by SDS-PAGE</b>	≥ 95% of 10 mg hIgG/1ml resin in 10mM Phosphate, 150 mM NaCl pH 7.4 at room temperature in 30 min.	pass
<b>Bioburden*</b>	0 cfu/vial	0 cfu/vial
<b>QC Release Date</b>		2018-11-30
<b>Expiration Date</b>		August 2019

\* Tested on Dextrose Agar, Fluid Thioglycollate Medium and Blood Agar.

**Linda Andersson**  
VP Production  
Genovis AB, Lund, SWEDEN