

Certificate of Analysis

Issued March 2026

Product Name	GlycINATOR® Immobilized Microspin
Product Number	A0-GL6-010, A0-GL6-025, A0-GL6-050
Product Lot Number	A0GL6-19B04G
Product Composition	Each Microspin column contains 0.1 ml settled resin of GlycINATOR enzyme immobilized on agarose. The resin is suspended in 20% ethanol.
Manufacturer	Genovis AB, Sweden
Storage Temperature	+4-8°C

TEST	SPECIFICATIONS	RESULTS LOT
Enzyme Activity by SDS-PAGE	0.1 ml settled resin hydrolyzes ≥95% N-linked Fc glycans of 0.5 mg human IgG1 when incubated in PBS (10 mM sodium phosphate, 140 mM NaCl, 2.7 mM KCl), pH 7.4, at room temperature for 15 minutes.	Pass
Bioburden*	Free from microbial contamination	Pass
Production Date		February 2026
Retest Date		February 2028

* Tested on Dextrose Agar, Fluid Thioglycollate Medium and Blood Agar.

Niloofar Mohiman

Niloofar Mohiman
Senior QC Scientist
Genovis AB, Sweden